



# NABL

## National Accreditation Board for Testing and Calibration Laboratories

(An Autonomous Body under Department of Science & Technology, Govt. of India)

### CERTIFICATE OF ACCREDITATION

## HI-TECH CALIBRATION SERVICES

has been assessed and accredited in accordance with the standard

**ISO/IEC 17025:2005**

"General Requirements for the Competence of Testing & Calibration Laboratories"

for its facilities at

No.209, VGP Nagar, Muggapair West, Chennai, Tamil Nadu

in the discipline of

**ELECTRO-TECHNICAL CALIBRATION**

(To see the scope of accreditation of this laboratory, you may also visit NABL website [www.nabl-india.org](http://www.nabl-india.org))

**Certificate Number** C-1262

**Issue Date** 07/11/2016



**Valid Until** 10/09/2017

This certificate remains valid for the Scope of Accreditation as specified in the annexure subject to continued satisfactory compliance to the above standard & the additional requirements of NABL.

Signed for and on behalf of NABL

Avijit Das  
Program Manager

Anil Relia  
Director

Prof. S. K. Joshi  
Chairman





# NABL

## SCOPE OF ACCREDITATION

Laboratory	Hi Tech Calibration Services, No. 209, 2nd Floor, VGP Nagar, Mugappair West, Chennai, Tamil Nadu	Issue Date	07.11.2016
Accreditation Standard	ISO/IEC 17025:2005	Valid Until	10.09.2017
Discipline	Electro-Technical Calibration	Page	1 of 6
Certificate Number	C-1262		
Last Amended on	-		

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability ( $\pm$ )	Remarks
<b>SOURCE</b>			
1. DC VOLTAGE <sup>#</sup>	0.1 mV to 329 mV 329 mV to 1000 V	1.2 % to 0.003 % 0.003 % to 0.002 %	Using Multiproduct Calibrator Fluke 5522A by Direct Method
2. AC VOLTAGE <sup>#</sup>	<b>45 Hz to 1 kHz</b> 3 mV to 30 mV 30 mV to 300 mV 300 mV to 300 V 300 V to 1000 V	0.3 % to 0.043 % 0.043 % to 0.02 % 0.02 % to 0.023 % 0.023 % to 0.04 %	Using Multiproduct Calibrator Fluke 5522A by Direct Method
3. DC CURRENT <sup>#</sup>	100 $\mu$ A to 329 $\mu$ A 329 $\mu$ A to 3.29 mA 3.29 mA to 329 mA 329 mA to 2.99 A 2.99 A to 20 A  20 A to 100 A 100 A to 1000 A	0.040 % to 0.026 % 0.026 % to 0.014 % 0.014 % to 0.013 % 0.013 % to 0.046 % 0.046 % to 0.12 %  2.1 % to 1.7 % 1.7 %	Using Multiproduct Calibrator Fluke 5522A by Direct Method  Using Multiproduct Calibrator with Current Coil Fluke 5522A & 5500A by Direct Method
4. AC CURRENT <sup>#</sup>	<b>45 Hz to 1 kHz</b> 33 $\mu$ A to 329 $\mu$ A 329 $\mu$ A to 3.29 mA 3.29 mA to 32.9 mA 32.9 mA to 329 mA 329 mA to 2.99 A 2.99 A to 20 A  <b>50 Hz</b> 20 A to 100 A 100 A to 1000 A	0.6 % to 0.2 % 0.2 % to 0.12 % 0.12 % to 0.06 % 0.06 % 0.06 % to 0.08 % 0.08 % to 0.2 %  3.1 % to 2.5 % 2.5 %	Using Multiproduct Calibrator Fluke 5522A by Direct Method  Using Multiproduct Calibrator with Current Coil Fluke 5522A & 5500A by Direct Method

Vishal Shukla  
Convenor

Avijit Das  
Program Manager







# NABL

## SCOPE OF ACCREDITATION

**Laboratory** Hi Tech Calibration Services, No. 209, 2nd Floor, VGP Nagar, Mugappair West, Chennai, Tamil Nadu  
**Accreditation Standard** ISO/IEC 17025:2005  
**Discipline** Electro-Technical Calibration **Issue Date** 07.11.2016  
**Certificate Number** C-1262 **Valid Until** 10.09.2017  
**Last Amended on** - **Page** 2 of 6

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability (±)	Remarks
5. DC RESISTANCE <sup>#</sup>	2 Ω to 30 Ω	0.6 % to 0.07 %	Using Multiproduct Calibrator Fluke 5522A by Direct Method
	30 Ω to 300 Ω	0.07 % to 0.011 %	
	300 Ω to 3 kΩ	0.011 %	
	3 kΩ to 300 kΩ	0.011 % to 0.008 %	
	300 kΩ to 3 MΩ	0.008 % to 0.02 %	
	3 MΩ to 30 MΩ	0.02 % to 0.04 %	
	30 MΩ to 290 MΩ	0.04 % to 0.4 %	
290 MΩ to 1.09 GΩ	0.4 % to 2 %		
6. DC RESISTANCE <sup>#</sup>	0.01 Ω to 0.1 Ω	3.89 % to 0.62 %	Using Decade Resistance Box & Decade Meg Ohm Box by Direct Method
	0.1 Ω to 1 Ω	0.62 % to 0.07 %	
	1 Ω to 10 Ω	0.07 % to 0.06 %	
	10 Ω to 100 kΩ	0.06 %	
	100 kΩ to 100 MΩ	0.06 % to 2.44 %	
	100 MΩ to 100 GΩ	2.44 % to 6.0 %	
7. FREQUENCY <sup>#</sup>	1 Hz to 1000 Hz	0.0011 % to 0.00051 %	Using Multiproduct Calibrator Fluke 5522A by Direct Method
	1000 Hz to 100 kHz	0.00051 % to 0.00065 %	
8. CAPACITANCE <sup>#</sup>	<b>1 kHz</b>		Using Multiproduct Calibrator Fluke 5522A by Direct Method
	0.35 nF to 0.6 nF	4.1 % to 2.6 %	
	0.6 nF to 1 nF	2.6 % to 1.8 %	
	<b>100 Hz</b>		
	1 nF to 7 nF	1.8 % to 0.5 %	
	7 nF to 300 nF	0.5 % to 0.4 %	
	<b>50 Hz</b>		
	300 nF to 30 μF	0.4 % to 0.6 %	
	30 μF to 33 mF	0.6 % to 1.7 %	
9. DC POWER <sup>#</sup>	1 V to 1000 V	0.11 % to 0.14 %	Using Multiproduct Calibrator Fluke 5522A by Direct Method
	1 A to 20 A		
	1 W to 20 kW		

  
 Vishal Shukla  
 Convenor

  
 Avijit Das  
 Program Manager



# NABL

## SCOPE OF ACCREDITATION

Laboratory	Hi Tech Calibration Services, No. 209, 2nd Floor, VGP Nagar, Mugappair West, Chennai, Tamil Nadu	Issue Date	07.11.2016
Accreditation Standard	ISO/IEC 17025:2005	Valid Until	10.09.2017
Discipline	Electro-Technical Calibration	Page	3 of 6
Certificate Number	C-1262		
Last Amended on	-		

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability ( $\pm$ )	Remarks
10. AC POWER 10 <sup>#</sup>	50 Hz @ UPF 60 V to 240 V 10 mA to 20 A 0.6 W to 4.8 kW	3.00 % to 0.07 %	Using Multiproduct Calibrator Fluke 5522A by Direct Method
11. POWER FACTOR 10 <sup>#</sup>	50 Hz @ 0.2 pF 60 V to 240 V 10 mA to 20 A 0.12 W to 0.96 kW	0.0021 to 0.0013	Using Multiproduct Calibrator Fluke 5522A by Direct Method
12. OSCILLOSCOPE <sup>#</sup> AMPLITUDE (Vertical axis Deflection Factor)	DC: 1.25 mV to 130 V @ 1 M $\Omega$ 1 kHz Square Wave @ 1 M $\Omega$ 10 mV to 130 V	0.75 % to 0.07 % 0.25 % to 0.15 %	Using Multiproduct Calibrator Fluke 5522A & SC600 by Direct Method
TIME BASE (Horizontal Axis Deflection Factor)	2 ns to 2 s	0.007 % to 0.2 %	
BANDWIDTH	Up to 500 MHz	5 %	

Vishal Shukla  
Convenor

Avijit Das  
Program Manager





# NABL

## SCOPE OF ACCREDITATION

Laboratory	Hi Tech Calibration Services, No. 209, 2nd Floor, VGP Nagar, Mugappair West, Chennai, Tamil Nadu		
Accreditation Standard	ISO/IEC 17025:2005		
Discipline	Electro-Technical Calibration	Issue Date	07.11.2016
Certificate Number	C-1262	Valid Until	10.09.2017
Last Amended on	-	Page	4 of 6

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability ( $\pm$ )	Remarks
<b>13. TEMPERATURE SIMULATION<sup>#</sup></b>			
RTD Pt 100	(-) 200 °C to 600 °C 600 °C to 800 °C	0.027 °C 0.18 °C	Using Multiproduct Calibrator Fluke 5522A by Direct Method
K Type Thermocouples	-200 °C to 1370 °C	0.23 °C	
J Type Thermocouples	-200 °C to 1150 °C	0.16 °C	
E Type Thermocouples	-200 °C to 950 °C	0.15 °C	
B Type Thermocouples	600 °C to 1820 °C	0.61 °C	
S Type Thermocouples	50 °C to 1760 °C	0.55 °C	
R Type Thermocouples	50 °C to 1760 °C	0.59 °C	
N Type Thermocouples	-100 °C to 1250 °C	0.18 °C	
T Type Thermocouples	-250 °C to 350 °C	0.5 °C	
<b>MEASURE</b>			
1. DC VOLTAGE <sup>#</sup>	0.1 mV to 100 mV 100 mV to 1 V 1 V to 1000 V	5 % to 0.01 % 0.01 % to 0.0054 % 0.0054 %	Using 6 ½ Digit Multimeter by Direct Method
2. AC VOLTAGE <sup>#</sup>	50 Hz to 1 kHz 0.1 V to 750 V	0.11 %	Using 6 ½ Digit Multimeter by Direct Method
3. DC CURRENT <sup>#</sup>	10 $\mu$ A to 100 $\mu$ A 100 $\mu$ A to 100 mA 100 mA to 3 A  3 A to 30 A	0.35 % to 0.09 % 0.09 % to 0.063 % 0.063 % to 0.2 %  0.36 %	Using 6 ½ Digit Multimeter by Direct Method  Using 6 ½ Digit Multimeter & Shunt by Direct Method

Vishal Shukla  
Convenor

Avijit Das  
Program Manager





# NABL

## SCOPE OF ACCREDITATION

Laboratory	Hi Tech Calibration Services, No. 209, 2nd Floor, VGP Nagar, Mugappair West, Chennai, Tamil Nadu		
Accreditation Standard	ISO/IEC 17025:2005		
Discipline	Electro-Technical Calibration	Issue Date	07.11.2016
Certificate Number	C-1262	Valid Until	10.09.2017
Last Amended on	-	Page	5 of 6

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability ( $\pm$ )	Remarks
4. AC CURRENT <sup>#</sup>	50 Hz to 1 kHz 100 $\mu$ A to 3 A	0.17 %	Using 6 ½ Digit Multimeter by Direct Method
	3 A to 30 A	0.46 %	Using 6 ½ Digit Multimeter & Shunt by Direct Method
5. DC RESISTANCE <sup>#</sup>	0.1 $\Omega$ to 10 $\Omega$	0.46 % to 0.06 %	Using 6 ½ Digit Multimeter by Direct Method
	10 $\Omega$ to 100 $\Omega$	0.06 % to 0.02 %	
	100 $\Omega$ to 1 M $\Omega$	0.02 % to 0.02 %	
	1 M $\Omega$ to 10 M $\Omega$	0.02 % to 0.05 %	
	10 M $\Omega$ to 100 M $\Omega$	0.05 % to 0.93 %	
100 M $\Omega$ to 1 G $\Omega$	0.93 % to 9.24 %		
6. FREQUENCY <sup>#</sup>	3 Hz to 10 Hz	0.08 % to 0.05 %	Using 6 ½ Digit Multimeter by Direct Method
	10 Hz to 100 Hz	0.05 % to 0.01 %	
	100 Hz to 300 kHz	0.01 %	
7. TIME <sup>#</sup> (Timer / Stop Watch)	1 s to 1000 s	0.002 s to 0.063 s	Using Time Totaliser by Direct Method
	1000 s to 5000 s	0.063 s to 0.25 s	
	5000 s to 86400 s	0.25 s to 1 s	
8. TEMPERATURE SIMULATION <sup>#</sup>			
RTD Pt 100	(-) 200 °C to 800 °C	0.20 °C	Using 6 ½ Digit Multimeter by Direct Method
K Type Thermocouples	(-)200 °C to 1370 °C	0.21 °C	Using Multiproduct Calibrator Fluke 5522A by Direct Method
J Type Thermocouples	(-)200 °C to 1200 °C	0.14 °C	
E Type Thermocouples	(-)200 °C to 1000 °C	0.12 °C	
B Type Thermocouples	600 °C to 1820 °C	0.6 °C	
S Type Thermocouples	50 °C to 1768 °C	0.58 °C	
R Type Thermocouples	50 °C to 1768 °C	0.59 °C	
N Type Thermocouples	(-)270 °C to 1300 °C	0.17 °C	
T Type Thermocouples	(-)100 °C to 400 °C	0.11 °C	

Vishal Shukla  
Convenor

Avijit Das  
Program Manager



# NABL

## SCOPE OF ACCREDITATION

Laboratory	Hi Tech Calibration Services, No. 209, 2nd Floor, VGP Nagar, Mugappair West, Chennai, Tamil Nadu	Issue Date	07.11.2016
Accreditation Standard	ISO/IEC 17025:2005	Valid Until	10.09.2017
Discipline	Electro-Technical Calibration	Page	6 of 6
Certificate Number	C-1262		
Last Amended on	-		

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability ( $\pm$ )	Remarks
9. DC HIGH VOLTAGE <sup>#</sup>	1 kV to 5 kV 5 kV to 30 kV	1.5 % 1.5 %	Using High Voltage Probe with DMM Fluke 80K - 40 / 80K - 6 & 87V by Direct Method
10. AC HIGH VOLTAGE <sup>#</sup>	50 Hz 1 kV to 5 kV 5 kV to 28 kV	2.5 % 2.5 % to 6.2 %	Using High Voltage Probe with DMM Fluke 80K - 40 / 80K - 6 & 87V by Direct Method

\* Measurement Capability is expressed as an uncertainty ( $\pm$ ) at a confidence probability of 95%

<sup>#</sup> The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

Vishal Shukla  
Convenor

Avijit Das  
Program Manager